



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), गुवाहाटी
All India Institute of Medical Sciences, Guwahati
स्वास्थ्य और परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक वैधानिक निकाय
(A statutory body under the aegis of Ministry of Health and Family Welfare, GoI)

File No. 4-146/2023-24/AIIMS/GHY/PROC-Forensic Medicine/2293

Date:25-01-2024

Corrigendum-01

Ref. GeM bid no. GEM/2023/B/4337269, Dated: 21-12-2023

Name of Equipment: - 64 Slide Computer Tomography

Amendment in Specifications

Sl. No.	Section/Point	Specification mentioned in the GeM bid document	Amended as
1	<u>Technical Specification</u> Spiral/Helical Section: Point No.5	Minimum slice thickness should be 0.5 mm and maximum 10 mm or better.	Minimum slice thickness should be 0.625 mm and maximum 10 mm or better.
2	<u>Additional Specification</u> Point No.7 Electrical Works	The supply line will be provided by the institute up to one point within the CT area	The supply line will be provided by the institute up to one point in UPS room with suitable switchgear.

All other points and conditions will be remained unchanged.

The bid submission date is hereby extended for another ten (10) days on GeM Portal.

Extended date may please be referred on Gem Portal against the Tender.

SD/-

(Administrative Officer)

AIIMS, Guwahati