



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), गुवाहाटी  
**All India Institute of Medical Sciences, Guwahati**  
स्वास्थ्य और परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक वैधानिक निकाय  
(A statutory body under the aegis of Ministry of Health and Family Welfare, GoI)

File No. 4-107/2023-24/AIIMS/GHY/PROC/ENT

Date: -25-10-2023

**Corrigendum-01**

Ref. GeM bid no. GEM/2023/B/3980948, Dated: 04-10-2023

Name of Equipment: - Trans Nasal Endoscopy System

**Amendment in Specifications**

Sl. No.	Specification mentioned in the GeM bid document	Amended as
1.	Complete System should include HD Flexible Video TNE scope with Video processor and inbuilt Light Xenon (150W) light source, Monitor, trolley and recording software.	Complete system should include HD flexible Video TNE scope with video processor and in-built light Xenon or LED light source, Monitor, trolley and recording software.
<b>Video Processor with built in Xenon light source (150 watt or more)</b>		
2.	Video Processor with built in Xenon Light Source (150 watt or more)	Video Processor with built in Xenon Light Source/LED CHIP on tip (150 watt or more)
3.	It should have single CCD color system.	It should have single CCD color system or support CHIP on TIP scope.
4.	It should have Xenon short arc lamp of 150 W with 3W LED as auxiliary lamp.	It should have Xenon short arc lamp of 150 Watt or LED light system with an option 3 W LED auxiliary lamp
<b>Trans nasal Endoscope- TNE scope (Esophagoscope)</b>		
5.	High-Definition true color CCD based video images	High-Definition true color CCD/ CMOS based Video images.
6.	Distal end diameter less than 5.5 mm.	No changes.
7.	Field of view- 140 degree or more	Field of view-120 degree or more
8.	Angulations U/D 210/120 and R/L- 120/120 degrees	Angulations- U/D-180/120 and above and R/L 120/120

P.T.O

**Physical Demonstration:** All the participant bidders must arrange physical demonstration of their quoted product before the Technical Evaluation Committee at the site of AIIMS, Guwahati. This physical demonstration is an integrated part of technical evaluation. Any bidder fails to arrange the above-mentioned physical demonstration as per the date and time schedule (which will be communicated to the bidders in due course of time) after opening the technical bids, his/her bid may not be considered and will be summarily rejected.

**Extension of Bid Submission Date: -** In effect of the above amendment in specifications the submission of technical bid is hereby extended up to 15 days.

Sd/-  
Procurement Cell  
AIIMS, Guwahati