



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), गुवाहाटी
All India Institute of Medical Sciences, Guwahati
स्वास्थ्य और परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक वैधानिक निकाय
(A statutory body under the aegis of Ministry of Health and Family Welfare, GoI)

File No. 4-111/2023-24/AIIMS/GHY/PROC-MS Office/1879

Date: 06-12-2023

Corrigendum-01

Ref. GeM bid no. GEM/2023/B/4259957, Dated: 29th November, 2023

Name of Equipment: - Pneumatic Tube System

Amendment in Specifications

Sl. No.	Specification mentioned in the GeM bid document	Amended as
1.	<u>Point no. 2. a.</u> System Safety Extra Low Voltage- The system must use safety extra low voltage (SELV- specified as 42 Volt DC throughout, except for main/AC power to controller, blowers, and power amplifiers.	<u>Point no. 2. a.</u> System Safety Extra Low Voltage- The system must use safety extra low voltage (SELV- specified as 25 to 50 Volt DC throughout, except for main/AC power to controller, blowers, and power amplifiers
2.	<u>Point no. 9 h.</u> Carrier size-380 mm	<u>Point no. 9 h.</u> Carrier size range 350-400mm
3.	<u>Point No. 13. d.</u> Cables Temperature range: stationary: +70 degree centigrade. Stationary: - 5 degree C to + 50 degree centigrade.	<u>Point No. 13. d.</u> Cables Temperature range: stationary: +70 degree centigrade. Stationary: - 5 degree C to + 80 degree centigrade.

Sd/-
(Administrative Officer)
AIIMS, Guwahati